



1



Attach the regulator to the cylinder

2



Attach the tygon tube to the regulator

3



Attach the gassing gap to the sensor and the tygon tube to the gassing gap

4



Turn the regulator's valve on and check the cylinder concentration

5



Check the gas concentration on the controller screen after 2 minutes

The Bump test is:

Successful when the value shown on the controller is similar to the cylinder concentration.

Unsuccessful if the value shown is significantly different from the cylinder concentration. Then the sensor requires calibration.